

Manganese Reagent Set, 0.006-0.700mg/L Mn, 25 mL

Številka izdelka: **2243300**

EUR Cena:
Na voljo

Kontakt

Nevarno



For the determination of Manganese with PAN Method. For 25 ml sample.

For determination of Manganese by the PAN Method. Hach Method 8149. Range: 0.006 - 0.700 mg/l Mn. Sample Size: 25 mL. Reagent Set includes: Ascorbic Acid Powder Pillows, Alkaline-Cyanide Reagent and PAN Indicator Solution (0.1%) for 100 tests.

Specifikacije

Ime metode:	PAN
Inštrument:	DR/3000, DR/2000
Merilno območje:	0,006 - 0,700 mg/L Mn
Posebna značilnost:	Manganese Reagent set, 25ml
Prostornina vzorca:	25 mL
Številka metode:	8149
Število testov:	100