

Chloride Reagent Set, 10-8000mg/L Cl

Številka izdelka: **2272600**

EUR Cena: Kontakt

Na voljo

Nevarno



For Mercuric Nitrate Digital Titrator Titration.

For determination of chloride by mercuric nitrate Digital Titrator titration. Hach Method 8206. Ranges: 10-160 and 100-8000 mg/l Cl-. Reagent Set includes: Diphenylcarbazone Reagent Powder Pillows, Mercuric Nitrate Digital Titrator Cartridges, 0.2256 N and 2.256 N. Approximately 40-100 Tests.

Specifikacije

Ime metode:	Mercuric Nitrate Titration
Merilno območje:	10 - 8000 mg/L Cl^-
Platforma:	Digitalni Titrator
Številka metode:	8206
Število testov:	100